ACORN

AUSTRALIAN COLLEGE OF
PERIOPERATIVE NURSES

| **JOURNAL OF PERIOPERATIVE NURSING**

Volume 29 | Issue 4

Article 3

12-1-2016

Factors affecting surgical plume evacuation compliance

Follow this and additional works at: <https://www.journal.acorn.org.au/jpn>

Part of the [Health Services Administration Commons](#), [Health Services Research Commons](#),
[Perioperative, Operating Room and Surgical Nursing Commons](#), and the [Surgery Commons](#)

This work is licensed under a [Creative Commons Attribution 4.0 License](#).

Recommended Citation

Holmes, Stephanie (2016) "Factors affecting surgical plume evacuation compliance," *Journal of Perioperative Nursing*: Vol. 29 : Iss. 4 , Article 3.

Available at: <https://doi.org/10.26550/2209-1092.1010>

<https://www.journal.acorn.org.au/jpn/vol29/iss4/3>

This Article is brought to you for free and open access by Journal of Perioperative Nursing. It has been accepted for inclusion in Journal of Perioperative Nursing by an authorized editor of Journal of Perioperative Nursing.

Factors affecting surgical plume evacuation compliance

Abstract

Despite the acknowledgement of the hazards of surgical plume, compliance with smoke evacuation is not routine.

This review examines the current literature on factors influencing compliance with smoke evacuation. Factors identified included the design of the smoke evacuation device, surgeon refusal, education and managerial support. Strong leadership, education and policy enforcement from a local facility level are required to improve surgical plume evacuation compliance. More research in this field would help to further strengthen these findings.

Key words: surgical plume, smoke evacuation, surgical smoke, compliance

Introduction

The Australian College of Operating Room Nurses (ACORN) standards and the recently released New South Wales (NSW) Health guideline include recommendations about surgical plume in the operating theatre; however, it is unclear what impact these have had on compliance. The evidence shows that surgical plume is dangerous to personnel yet, from experience, compliance with smoke evacuation is less than ideal. The objective of this review is to identify and appraise the best available evidence on factors that influence compliance with surgical plume evacuation.

Surgical plume is generated by heat-generating devices such as electrosurgical units, lasers, ultrasonic devices, high-speed drills, burrs and saws¹. These pieces of

equipment are vital and are used in many surgical procedures worldwide. Surgical plume, also called surgical smoke, is the result of thermal destruction of bone or tissue². In the United States, it is estimated that more than 500 000 health care workers are exposed to surgical plume every year³. Potential risks to health care workers exposed to surgical plume include acute and chronic inflammatory respiratory changes, eye irritation and headache¹.

Toxic substances, pathogens, mutagens and carcinogens are released into the atmosphere by electrosurgery, powered instruments and lasers². A systematic review conducted by Mowbray et al.⁴ confirmed that surgical plume contains potentially carcinogenic compounds physically small enough to inhale and reach the lower airways. Surgical plume can contain a variety of contaminants, including bacteria, viruses, cellular debris, gases, vapours and fumes⁵. Each heat-generating device produces particles of a different size¹. The smaller the particle size, the further it travels¹. This means that all personnel in the operating theatre can be affected, not just personnel in the surgical field. Surgical plume can also affect patients. During laparoscopic surgery, surgical plume is absorbed into the peritoneal cavity⁶. A study conducted by Beebe et al.⁷ to determine detectable levels of carbon monoxide produced by electrosurgery during laparoscopic cholecystectomy procedures revealed that carbon monoxide was present in the peritoneal cavity within five minutes. Both intraoperative and postoperative carbon monoxide levels exceeded levels recommended

Author

Stephanie Holmes

MN

Nurse Unit Manager, Operating Theatres
Wollongong Private Hospital

by the Occupational Safety and Health Administration⁷.

Standards Australia stipulate that when laser is used contaminants need to be evacuated appropriately from the surgical field⁸. Airborne contaminants need to be captured as close as possible to the point of evolution and removed by localised exhaust ventilation (LEV)⁸. Compared to laser, less attention has been paid to mitigating the risks associated with exposure to diathermy plume. However, in 2015 NSW Health released a guideline 'Work health and safety – Controlling exposure to surgical plume'. This guideline states that each NSW health organisation has a primary duty to ensure the health and safety of workers and other persons in the workplace⁵. The NSW health organisation must eliminate risks to health and safety so far as is reasonably practicable and if eliminating the risk is not practicable, to minimise the risk⁵. Control measures include adequate plume evacuation at the source. Plume evacuation systems aim to remove plume from the environment, reducing potential hazards of exposure⁵. The ACORN standard 'Surgical plume' states exposure to surgical plume is a workplace health and safety hazard and must be mitigated by methods appropriate to the procedure and instrumentation⁹.

Smoke evacuation systems vary. Many are portable and can be activated by a foot switch or automatically⁵. These systems capture smoke through tubing positioned usually within five centimetres of the plume source⁵. The captured plume is passed through filters and the filtered air is

then released back into the room⁵. These plume evacuation systems need to have ultra-low penetrating air (ULPA) filters with an efficiency rating of not less than 99.999 per cent⁹.

For endoscopic or laparoscopic procedures, both active and passive devices exist. They are single use devices that provide improved visibility into the peritoneal cavity without compromising room air or pneumoperitoneum and minimise patient exposure to surgical plume⁵.

Local stationary evacuation systems have a suction source located in a surgical boom or in the interstitial space above the operating theatre⁵. These systems capture plume with a capture device and hose and vent the filtered air inside or outside of the theatre⁵. If medical vacuum systems are used, an in-line ULPA filter must be used between the fluid trap and the vacuum regulator⁵. Filters are to be changed according to manufacturer's instructions and hospital policy⁵.

Despite the evidence showing the dangers of surgical plume, research shows that compliance is still low. In a study carried out in the United Kingdom, a list of 56 British plastic surgery units was obtained from the British Association of Plastic Reconstructive and Aesthetic Surgeons (BAPRAS)¹⁰. Of the 56 plastic surgery units, 50 responded to the survey. Smoke evacuators were available for use in 66 percent of plastic surgery units. Despite this, their use was not universal and was dependent upon the surgeon and the procedure¹⁰.

The Search

An initial, limited search of MEDLINE and Cumulative Index to Nursing and Allied Health Literature (CINAHL) was undertaken followed by analysis of the text words contained in the

title and abstract and of the index terms used to describe the article. A second search was undertaken using all identified keywords and index terms. Keywords included surgical smoke, surgical plume, smoke evacuation, compliance and barriers. The reference lists of all identified reports and articles were searched for additional studies. All studies published in English were considered. As surgical plume is produced by laser and electrosurgery, all studies discussing these methods were included. Studies that did not discuss compliance or factors influencing compliance were excluded. A total of fifteen articles were used in this literature review.

Results

Results were varied and produced multiple types of research: literature reviews, cross-sectional surveys and many questionnaires. Results of the literature search have been separated into themes.

Surgeon refusal

According to a study conducted by Shultz¹¹ a dismissive attitude towards smoke inhalation is often the decisive factor in the choice not to use smoke evacuation devices. Similarly, in a random sample study conducted by Ball¹² surgeon refusal was a common barrier to smoke evacuation compliance. A literature review by Lindsey et al.¹³ reported nurses did not feel empowered to use protective equipment because this use was at the discretion of the surgeons. In a survey to identify compliance with smoke evacuation, carried out by Edwards and Reiman¹⁴, it was found that the most commonly reported obstacle was surgeon resistance or refusal. Similar results were found in a cross-sectional survey in the United Kingdom. In this survey, three per cent of surgeons used dedicated smoke evacuators,

despite the fact that 72 per cent felt that inadequate precautions were taken to protect staff and patients¹⁵.

Smoke evacuation system design

Edwards and Reiman¹⁴ found that obstacles to the use of smoke evacuator systems included the bulkiness of the systems which caused them to get in the way and take up too much room. Other obstacles found by Edwards and Reiman¹⁴ included excessive noise produced by the smoke evacuator system. Ball¹² also found that excessive noise was a deterrent to compliance. This was partly due to the fact that older models were in use, which can be noisier than newer models on the market today¹⁴. Newer models of smoke evacuators are padded to abate the noise produced when the system is activated¹². The bulkiness of the smoke evacuation pencil was also mentioned as a barrier¹⁴. The handpieces were found to be too heavy, awkward and prone to clogging¹⁴.

Attitudes of personnel

In an analysis of surgical plume capture and evacuation, Shultz¹¹ found a dismissive attitude towards the risks of plume is often the decisive factor in the choice not to use smoke evacuation devices. Similarly, in a web-based survey, several respondents stated that physicians did not recognise surgical plume as a hazard¹⁴. A cross-sectional survey completed by consultant surgeons, registrars and perioperative nurses on current attitudes and practices towards diathermy smoke found registrars (70 per cent) were more likely to use evacuation equipment than consultants (43 per cent)¹⁵. There was uncertainty amongst the consultants as to the dangers of surgical plume, and a belief that more evidence was

required¹⁵. Ball³ conducted a random sample survey of perioperative nurses and concluded that physicians need to be educated on the documented hazards of smoke inhalation. However, results from Ball³ also showed that if nurses' perceptions about surgical plume recommendations were positive, they were more likely to comply with recommendations.

Managerial support

A web-based survey targeted at perioperative nurses found a lack of support from management contributed to non-compliance¹⁶. A cross-sectional survey interviewing perioperative nurses in the United States found that strong leadership support is a key component to compliance³. Ball³ also found that leaders must show a keen interest in making sure equipment is available and mandating use through policy enforcement. Scott et al.¹⁷ carried out a quality project in their operating theatre suite to improve compliance with smoke evacuation. It was recognised that management played an important role in supporting the nursing staff to use evacuation equipment and rewarding changes in practice¹⁷.

Education

A random sample survey reported that education and demonstration of equipment helps to convince nurses of the need to evacuate surgical plume³. The study also found if smoke evacuation policies are easy to understand and implement, nurses will comply with them³. Education and training programs are important for compliance: if nurses have knowledge about the hazards of surgical plume, they are more likely to want to use smoke evacuators¹². A literature review by Lindsey et al.¹³ on the hazards of diathermy plume found that nurses used lower

levels of protection with diathermy compared to laser. Surgeons also need to receive education about the hazards of surgical plume¹². Ball¹² found that many surgeons need to see the evidence before they are willing to change their practice. Edwards and Reiman¹⁶ found that the frequency of smoke evacuation reflected the clinicians' perception of the relative hazard.

Discussion

This literature review shows that there are various factors influencing smoke evacuation compliance. Limitations of this literature review include the quality of evidence: the majority of studies were cross-sectional surveys, many with a small sample size. Response rates in many studies were also low. The literature mainly focused on perioperative nurses' perceptions. Whilst perioperative nurses' perceptions are important, perhaps more future studies could include surgeon perceptions as the literature demonstrates surgeon refusal is a barrier to successful smoke evacuation compliance. Many studies were also online. Perhaps further studies could be observational and onsite to determine factors influencing compliance, as opposed to perception. Future studies could compare hospitals to determine significant differences in compliance. The majority of studies included in this literature review were conducted in the United States and the United Kingdom; research in Australia, particularly in NSW, to determine the impact of the NSW Health guideline would be beneficial.

Conclusion

This literature review has shown there are many factors, both positive and negative, influencing compliance with smoke evacuation. Strong leadership and education are

vital to ensuring smoke evacuation compliance. Managers need to agree and stand strong on smoke evacuation. Managers generally tend to have positive relationships with surgeons, which is important to work through the barrier of surgeon refusal¹⁸. Physicians that support smoke evacuation should be supported and encouraged to serve as advocates¹⁴. As stated in the NSW Health guideline, policies should be developed at a local level and they should be developed in consultation with surgeons. Policies should be clear, concise and simple to follow. Policies should state when evacuation devices are required to be used and represent the department's stance on plume evacuation. Education, not just for nursing staff but for all perioperative personnel, is important. Surgeons need to be made aware of the hazards of surgical plume and the impact this could potentially have on personnel within the theatre. The literature shows that when nurses are aware of the hazards, compliance increases³. This could potentially be the same with surgeons. Education programs should be implemented in all operating suites and made mandatory. Smoke evacuation practices should be monitored regularly for compliance. Compliance should be recognised and non-compliance should be investigated to determine the possible cause.

References

1. Ulmer BC. The hazards of surgical smoke. *AORN J* 2008;87(40):721–738.
2. Marsh S. The smoke factor: Things you should know. *J Perioper Pract* 2012;22(3):91–94.
3. Ball K. Surgical smoke evacuation guidelines: Compliance among perioperative nurses. *AORN J* 2010;92(2):e1–e23.
4. Mowbray N, Ansell J, Warren N, Wall P, Torkington J. Is surgical smoke harmful to theater staff? A systematic review. *Surg Endosc* 2013;27:3100–3107.

5. Work Health and Safety – Controlling exposure to surgical plume [Internet]. Sydney: Ministry of Health, NSW; 2015 Jan 19 [cited 2015 Jul 21]. Available from: www.health.nsw.gov.au/policies/gl/2015/pdf/GL2015_002.pdf.
6. Watson DS. Surgical smoke evacuation during laparoscopic surgery. *AORN J* 2010;92(3):347–350.
7. Beebe SS, Swica H, Carlson N, Palahniuk RJ and Goodale RL. High levels of carbon monoxide are produced by electrocautery of tissue during laparoscopic cholecystectomy. *Anesth Analg* 1993; 77(2):338–341.
8. Guide to the safe use of lasers in health care (AS/NZ 4173:2004). Standards Australia and Standards New Zealand 2004.
9. ACORN Standards for Perioperative Nursing 2014–2015. Adelaide, South Australia: The Australian College of Operating Room Nurses Ltd, 2014.
10. Hill DS, O'Neill JK, Powell RJ and Oliver DW. Surgical smoke – a health hazard in the operating theatre: A study to quantify exposure and a survey of the use of smoke evacuator systems in the UK plastic surgery units. *J Plast Reconstr Aesthet Surg* 2012;65:911–916.
11. Schultz L. An analysis of surgical smoke plume components, capture and evacuation. *AORN J* 2014;99(2):289–298.
12. Ball K. Compliance with surgical smoke evacuation guidelines: Implications for practice. *AORN J* 2010;92(2):142–149.
13. Lindsey C, Hutchinson M, Mellor G. The nature and hazards of diathermy plumes: A review. *AORN Journal* 2015;101(4):428–442.
14. Edwards BE, Reiman RE. Comparison of current and past surgical smoke control practices. *AORN J* 2012;95(3):337–350.
15. Spearman J, Tsavellas G, Nichols P. Current attitudes and practices towards diathermy smoke. *Ann R Coll Surg Engl* 2007;89:162–165.
16. Edwards BE, Reiman RE. Results of a survey on current surgical smoke control practices. *AORN J* 2008;87(4):739–749.
17. Scott H, Mustard P, Cooper H, Hayde C. Development of a plume evacuation policy: A health and safety issue. *The Dissector* 2014;41(4):10–14.
18. Ball K. Surgical smoke evacuation guidelines: Are you compliant? *OR Nurse J* 2011;5(6):40–44.

Credentialing for health professionals

Credentialing is part of a wider organisational quality and risk-management system designed primarily to protect patients.

Credentialing has the potential to improve safety for patients by ensuring clinicians practise within the bounds of their training and competency, and within the capacity of the service in which they are working.

'Credentialing' refers to the formal process used to verify the qualifications, experience, professional standing and other relevant professional attributes of medical practitioners for the purpose of forming a view about their competence, performance and professional suitability to provide safe, high quality health care services within specific organisational environments. These processes are generally mandatory for specialist medical

practitioners in the public and private hospital sector.

Credentialing by health services has largely focused on specialist medical practitioners, but it has the potential for wider application to other health professions.

The former Australian Council for Safety and Quality in Health Care developed the 'Standard for credentialing and defining the scope of clinical practice of medical practitioners, for use in public and private hospitals' in 2004. Its implementation is underway in all jurisdictions and across the private hospital sector. The structures and processes being used vary between states and different health care settings.

The information above is from The Australian Commission on Safety and Quality in Health Care website at <https://www.safetyandquality.gov.au/our-work/credentialing>.

ACORN AUSTRALIAN COLLEGE OF PERIOPERATIVE NURSES

Would you like to be a credentialed specialist nurse?

ACORN is conducting a survey of its members to gauge the level of interest in credentialing.

While there are many advantages, the process of credentialing requires administration and regulation. It can be time consuming and has an associated economic cost.

ACORN would like to know what you think about credentialing. We have developed a survey at <https://www.surveymonkey.com/r/CREDENT>. It only takes a few minutes and is completely anonymous.

Please take part.